

**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Laboratory Name      Central Laboratory (Thailand) Company Limited  
 Address                50 Phaholyothin Road, Ladayao, Jatujak, Bangkok  
 Accreditation no.     CALIBRATION 0125  
 Laboratory Status     Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
1. Temperature	Liquid in glass thermometer Total immersion -40 °C to 250 °C  Temperature indicator/logger with sensor Resistance thermometer -40 °C to 250 °C Thermocouple Type E, J, K, T and N -40 °C to 50 °C > 50 °C to 100 °C > 100 °C to 150 °C > 150 °C to 200 °C > 200 °C to 250 °C > 250 °C to 300 °C > 300 °C to 400 °C > 400 °C to 600 °C  Type S and R 200 °C to 650 °C  Dial thermometer -40 °C to 250 °C	0.044 °C  0.034 °C  0.14 °C 0.25 °C 0.36 °C 0.47 °C 0.59 °C 1.5 °C 1.8 °C 2.3 °C  1.7 °C  0.59 °C	In-house method : WI-TMP-105-CC based on ASTM E 77  In-house method : WI-TMP-114-CC by comparison technique  In-house method : WI-TMP-111-CC by comparison technique

\* Expressed as an uncertainty ( $\pm$ ) providing a level of confidence of approximately 95 %

**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Accreditation no.

CALIBRATION 0125

Laboratory Status

Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
1. Temperature (cont.)	<p>Temperature graph recorder with sensor</p> <p>Resistance thermometer 0 °C to 150 °C</p> <p>PRT and IPRT 4 wire sensor -40 °C to 250 °C</p> <p>Thermocouple sensor Type J, K and T 0 °C to 100 °C &gt; 100 °C to 200 °C</p> <p>Thermo-hygrometer, Temperature &amp; humidity data logger</p> <p>Temperature 10 °C to 60 °C</p> <p>Relative humidity @ 25 °C 15 % to 30 % &gt; 30 % to 70 % &gt; 70 % to 90 %</p>	<p>0.60 °C</p> <p>0.025 °C</p> <p>0.40 °C 0.75 °C</p> <p>0.42 °C</p> <p>2.4 % 2.2 % 2.4 %</p>	<p>In-house method : WI-TMP-114-CC by comparison technique</p> <p>In-house method : WI-TMP-116-CC based on ASTM E 644-04, BS EN 60751 : 1996 and IEC 751 : 1983</p> <p>In-house method : WI-TMP-117-CC based on ASTM E 220/E 230</p> <p>In-house method : WI-HUM-001-CC by comparison with temperature humidity meter in controlled chamber</p>

\* Expressed as an uncertainty ( $\pm$ ) providing a level of confidence of approximately 95 %

**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Accreditation no. **CALIBRATION 0125**

Laboratory Status  Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
2. Chemical	Burette 2.5 cm <sup>3</sup> > 2.5 cm <sup>3</sup> to 10 cm <sup>3</sup> > 10 cm <sup>3</sup> to 25 cm <sup>3</sup> > 25 cm <sup>3</sup> to 50 cm <sup>3</sup> > 50 cm <sup>3</sup> to 100 cm <sup>3</sup>  Volumetric flask 5 cm <sup>3</sup> 10 cm <sup>3</sup> 20 cm <sup>3</sup> 25 cm <sup>3</sup> 50 cm <sup>3</sup> 100 cm <sup>3</sup> 200 cm <sup>3</sup> 250 cm <sup>3</sup> 500 cm <sup>3</sup> 1 000 cm <sup>3</sup> 2 000 cm <sup>3</sup>  Volumetric pipette 0.5 cm <sup>3</sup> 1 cm <sup>3</sup> 2 cm <sup>3</sup> 3 cm <sup>3</sup> 4 cm <sup>3</sup> 5 cm <sup>3</sup> 10 cm <sup>3</sup> 15 cm <sup>3</sup> 20 cm <sup>3</sup>	0.003 6 cm <sup>3</sup> 0.003 8 cm <sup>3</sup> 0.006 7 cm <sup>3</sup> 0.011 cm <sup>3</sup> 0.022 cm <sup>3</sup>  0.005 9 cm <sup>3</sup> 0.006 0 cm <sup>3</sup> 0.006 4 cm <sup>3</sup> 0.006 6 cm <sup>3</sup> 0.010 cm <sup>3</sup> 0.018 cm <sup>3</sup> 0.029 cm <sup>3</sup> 0.036 cm <sup>3</sup> 0.064 cm <sup>3</sup> 0.13 cm <sup>3</sup> 0.26 cm <sup>3</sup>  0.002 4 cm <sup>3</sup> 0.002 4 cm <sup>3</sup> 0.002 4 cm <sup>3</sup> 0.002 5 cm <sup>3</sup> 0.002 5 cm <sup>3</sup> 0.002 5 cm <sup>3</sup> 0.004 5 cm <sup>3</sup> 0.006 4 cm <sup>3</sup> 0.006 5 cm <sup>3</sup>	In-house method : WI-VOL-101-CC based on ASTM E 542-2001

\* Expressed as an uncertainty ( $\pm$ ) providing a level of confidence of approximately 95 %

**Scope of Accreditation for Calibration**

Certificate No. 20C110/0789

Accreditation no. CALIBRATION 0125

Laboratory Status  Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
2. Chemical (cont.)	Volumetric pipette (cont.)		In-house method : WI-VOL-101-CC based on ASTM E 542-2001
	25 cm <sup>3</sup>	0.007 3 cm <sup>3</sup>	
	50 cm <sup>3</sup>	0.012 cm <sup>3</sup>	
	100 cm <sup>3</sup>	0.018 cm <sup>3</sup>	
	Graduated pipette		
	0.5 cm <sup>3</sup>	0.002 4 cm <sup>3</sup>	
	> 0.5 cm <sup>3</sup> to 1 cm <sup>3</sup>	0.002 4 cm <sup>3</sup>	
	> 1 cm <sup>3</sup> to 5 cm <sup>3</sup>	0.002 5 cm <sup>3</sup>	
	> 5 cm <sup>3</sup> to 10 cm <sup>3</sup>	0.004 3 cm <sup>3</sup>	
	> 10 cm <sup>3</sup> to 25 cm <sup>3</sup>	0.009 9 cm <sup>3</sup>	
	Piston pipette		In-house method : WI-VOL-102-CC based on ISO 8655 : 2002
	> 10 µl to 50 µl	0.069 µl	
	> 50 µl to 100 µl	0.19 µl	
	> 100 µl to 500 µl	0.23 µl	
	> 500 µl to 1 000 µl	0.30 µl	
	> 1 000 µl to 2 500 µl	1.3 µl	
	> 2 500 µl to 5 000 µl	1.4 µl	
	> 5 000 µl to 10 000 µl	1.8 µl	
	Cylinder		In-house method : WI-VOL-101-CC based on ASTM E 542-2001
	5 cm <sup>3</sup>	0.009 6 cm <sup>3</sup>	
	> 5 cm <sup>3</sup> to 10 cm <sup>3</sup>	0.009 7 cm <sup>3</sup>	
	> 10 cm <sup>3</sup> to 25 cm <sup>3</sup>	0.016 cm <sup>3</sup>	
	> 25 cm <sup>3</sup> to 50 cm <sup>3</sup>	0.020 cm <sup>3</sup>	
	> 50 cm <sup>3</sup> to 100 cm <sup>3</sup>	0.026 cm <sup>3</sup>	
	> 100 cm <sup>3</sup> to 250 cm <sup>3</sup>	0.049 cm <sup>3</sup>	
	> 250 cm <sup>3</sup> to 500 cm <sup>3</sup>	0.079 cm <sup>3</sup>	
	> 500 cm <sup>3</sup> to 1 000 cm <sup>3</sup>	0.14 cm <sup>3</sup>	

\* Expressed as an uncertainty (±) providing a level of confidence of approximately 95 %

**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Accreditation no. **CALIBRATION 0125**

Laboratory Status  Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
3. Mass	Conventional mass Class F1		OIML R 111-1
	1 g	0.019 mg	
	2 g	0.022 mg	
	5 g	0.027 mg	
	10 g	0.033 mg	
	20 g	0.040 mg	
	50 g	0.048 mg	
	100 g	0.078 mg	
	200 g	0.16 mg	
	500 g	0.40 mg	
	1 kg	1.3 mg	
	2 kg	1.9 mg	
	5 kg	4.0 mg	
	Class M1		
	0.1 g	0.013 mg	
	0.2 g	0.015 mg	
	0.5 g	0.019 mg	
	10 kg	0.092 g	
	20 kg	0.11 g	
	Conventional mass		
	0.01 g	0.011 mg	
	0.02 g	0.012 mg	
	0.05 g	0.012 mg	
	0.1 g	0.013 mg	
	0.2 g	0.014 mg	
	0.5 g	0.016 mg	
	1 g	0.019 mg	

\* Expressed as an uncertainty ( $\pm$ ) providing a level of confidence of approximately 95 %

**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Accreditation no. **CALIBRATION 0125**

Laboratory Status  Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
3. Mass (cont.)	Conventional mass (cont.)		OIML R 111-1
	2 g	0.022 mg	
	5 g	0.027 mg	
	10 g	0.033 mg	
	20 g	0.040 mg	
	50 g	0.048 mg	
	100 g	0.078 mg	
	200 g	0.16 mg	
	500 g	0.40 mg	
	1 kg	1.3 mg	
	2 kg	1.9 mg	
	5 kg	4.0 mg	
	10 kg	0.11 g	
	20 kg	0.12 g	
4. Dimension	Micrometer caliper for external measurement		In-house method : WI-DMS-101-CC based on JIS B 7502 : 1994
	0 mm to 15 mm	0.95 µm	
	> 15 mm to 25 mm	1.1 µm	
	Vernier, dial and digital caliper		In-house method : WI-DMS-102-CC based on JIS B 7507 : 1993
	0 mm to 300 mm	16 µm	

\* Expressed as an uncertainty ( $\pm$ ) providing a level of confidence of approximately 95 %

**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Accreditation no. **CALIBRATION 0125**

Laboratory Status  Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
1. Temperature	Temperature indicator/logger with sensor Resistance thermometer -30 °C to 125 °C > 125 °C to 300 °C Thermocouple Type E, J, K, T and N -30 °C to 50 °C > 50 °C to 100 °C > 100 °C to 125 °C > 125 °C to 200 °C > 200 °C to 250 °C > 250 °C to 300 °C  Temperature graph recorder with sensor Resistance thermometer 0 °C to 150 °C  Dial thermometer -30 °C to 125 °C  Liquid bath 0 °C to 50 °C > 50 °C to 100 °C  Autoclave 110 °C to 125 °C	0.13 °C 1.5 °C  0.17 °C 0.26 °C 0.37 °C 0.52 °C 0.60 °C 0.71 °C  0.60 °C  0.59 °C  0.13 °C 0.17 °C  0.76 °C	In-house method : WI-TMP-114-CC by comparison technique  In-house method : WI-TMP-114-CC by comparison technique  In-house method : WI-TMP-111-CC by comparison technique  In-house method : WI-TMP-001-CC based on ASTM E 715-80 : 2001  In-house method : WI-TMP-003-CC based on BS 2646-5 : 1993

\* Expressed as an uncertainty ( $\pm$ ) providing a level of confidence of approximately 95 %

**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Accreditation no. **CALIBRATION 0125**

Laboratory Status  Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
1. Temperature (cont.)	Temperature controlled chamber  -40 °C to 0 °C > 0 °C to 70 °C > 70 °C to 150 °C > 150 °C to 200 °C	1.1 °C 0.47 °C 0.79 °C 1.3 °C	In-house method : WI-TMP-002-CC based on ASTM E 145-94 : 2001
	Furnace  200 °C to 650 °C	2.7 °C	In-house method : WI-TMP-004-CC by comparison technique
2. Mass	Electronic balance  1 mg to 10 mg > 10 mg to 50 mg > 50 mg to 100 mg > 100 mg to 500 mg > 500 mg to 1 000 mg > 1 g to 5 g > 5 g to 10 g > 10 g to 20 g > 20 g to 40 g > 40 g to 60 g > 60 g to 80 g > 80 g to 100 g > 100 g to 120 g > 120 g to 140 g > 140 g to 160 g > 160 g to 200 g > 200 g to 300 g	9.4µg 10 µg 11µg 12 µg 14 µg 21 µg 26 µg 37 µg 73 µg 88 µg 0.13 mg 0.14 mg 0.17 mg 0.21 mg 0.23 mg 0.26 mg 0.41 mg	In-house method : WI-MAS-002-CC based on UKAS LAB 14 : 2015

\* Expressed as an uncertainty (±) providing a level of confidence of approximately 95 % 

**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Accreditation no. **CALIBRATION 0125**

Laboratory Status  Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
2. Mass (cont.)	Electronic balance (cont.) > 300 g to 500 g > 500 g to 1 000 g > 1 000 g to 1 500 g > 1 500 g to 2 000 g > 2 000 g to 2 500 g > 2 500 g to 3 000 g > 3 000 g to 4 000 g > 4 000 g to 5 000 g > 5 kg to 10 kg > 10 kg to 20 kg > 20 kg to 30 kg > 30 kg to 60 kg > 60 kg to 120 kg > 120 kg to 150 kg	0.66 mg 1.6 mg 2.2 mg 2.8 mg 3.4 mg 3.9 mg 6.5 mg 7.0 mg 0.085 g 0.091 g 0.20 g 0.84 g 8.4 g 8.6 g	In-house method : WI-MAS-002-CC based on UKAS LAB 14 : 2015
3. Electrical	Centrifuge Shaker or rotator 50 r/min to < 1 000 r/min 1 000 r/min to 15 000 r/min	0.61 r/min 1.5 r/min	In-house method : WI-SPD-001-CC by direct measurement with digital tachometer
4. Chemical	Breath alcohol tester 45 mg/100 ml to 55 mg/100 ml	2.2 mg/100 ml	In-house method : WI-CHM-002-CC by wet bath simulator technique

\* Expressed as an uncertainty ( $\pm$ ) providing a level of confidence of approximately 95 %



**Scope of Accreditation for Calibration**

**Certificate No. 20C110/0789**

Accreditation no. **CALIBRATION 0125**

Laboratory Status  Permanent  Site  Temporary  Mobile

Field of Calibration	Parameter	Calibration and Measurement Capability*	Method Used
4. Chemical (cont.)	pH meter		In-house method : WI-CHM-001-CC based on ASTM E 70-07
	DC Voltage		
	-177.48 mV	0.068 mV	
	-128.97 mV	0.068 mV	
	0 mV	0.068 mV	
	8.28 mV	0.068 mV	
	177.48 mV	0.068 mV	
	314.73 mV	0.068 mV	
	Nominal pH		
	1.7	0.007 8	
	4	0.007 8	
	7	0.007 8	
	9	0.007 8	
	10	0.007 8	

\* Expressed as an uncertainty ( $\pm$ ) providing a level of confidence of approximately 95 % 

Issue date : 23<sup>th</sup> September B.E. 2563 (2019)

(Signature)

(Mr. Verakit Rantakittanawat)

Deputy Secretary-General

for Secretary-General, Thai Industrial Standards Institute